

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092724\
 Data File : VN084207.D
 Acq On : 27 Sep 2024 16:42
 Operator : JC\MD
 Sample : P4190-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-T

Quant Time: Sep 28 01:01:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 09/30/2024
 Supervised By :Mahesh Dadoda 09/30/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	170068	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	303129	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	256636	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	100086	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	124862	48.691	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	97.380%		
35) Dibromofluoromethane	8.171	113	94596	48.746	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.500%		
50) Toluene-d8	10.565	98	361124	53.829	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	107.660%		
62) 4-Bromofluorobenzene	12.847	95	116737	42.176	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	84.360%		
Target Compounds						
16) Acetone	4.442	43	25780	25.078	ug/l	95
20) Methylene Chloride	5.277	84	3341	1.588	ug/l #	88
43) Isopropyl Acetate	8.688	43	217883m	45.350	ug/l	
95) Naphthalene	15.641	128	26445	4.771	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092724\
Data File : VN084207.D
Acq On : 27 Sep 2024 16:42
Operator : JC\MD
Sample : P4190-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-T

Quant Time: Sep 28 01:01:02 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 09/30/2024
Supervised By :Mahesh Dadoda 09/30/2024

