

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092724\
 Data File : VN084208.D
 Acq On : 27 Sep 2024 17:06
 Operator : JC\MD
 Sample : P4192-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VNJ-222

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 09/30/2024
 Supervised By :Mahesh Dadoda 09/30/2024

Quant Time: Sep 28 01:01:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	169159	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	303215	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	260855	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	105878	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	125325	49.134	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	98.260%		
35) Dibromofluoromethane	8.171	113	94810	48.842	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.680%		
50) Toluene-d8	10.565	98	360325	53.695	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	107.380%		
62) 4-Bromofluorobenzene	12.847	95	121117	43.746	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	87.500%		
Target Compounds						
16) Acetone	4.436	43	29220	28.577	ug/l	98
20) Methylene Chloride	5.283	84	3540	1.692	ug/l #	86
43) Isopropyl Acetate	8.688	43	147820m	30.758	ug/l	
95) Naphthalene	15.641	128	13969	2.383	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092724\
Data File : VN084208.D
Acq On : 27 Sep 2024 17:06
Operator : JC\MD
Sample : P4192-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VNJ-222

Quant Time: Sep 28 01:01:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 09/30/2024
Supervised By :Mahesh Dadoda 09/30/2024

