

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092820\
 Data File : VN063700.D
 Acq On : 28 Sep 2020 13:04
 Operator : JC/MD
 Sample : VN0928WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0928WBL01

Quant Time: Sep 29 03:51:55 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	212242	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	388656	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	396229	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	165206	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.99	65	185124	51.24	ug/l	0.00
Spiked Amount			Recovery	=	102.48%	
35) Dibromofluoromethane	7.55	113	132995	49.43	ug/l	0.00
Spiked Amount			Recovery	=	98.86%	
50) Toluene-d8	10.06	98	509224	50.33	ug/l	0.00
Spiked Amount			Recovery	=	100.66%	
62) 4-Bromofluorobenzene	12.38	95	194625	52.48	ug/l	0.00
Spiked Amount			Recovery	=	104.96%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092820\
Data File : VN063700.D
Acq On : 28 Sep 2020 13:04
Operator : JC/MD
Sample : VN0928WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0928WBL01

Quant Time: Sep 29 03:51:55 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

