

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092820\
 Data File : VN063705.D
 Acq On : 28 Sep 2020 15:09
 Operator : JC/MD
 Sample : L4133-02DL 10X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW1DDL

Quant Time: Sep 29 04:08:22 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	212768	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	393624	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	421292	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	186015	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	183660	50.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.42%	
35) Dibromofluoromethane	7.55	113	133784	49.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.20%	
50) Toluene-d8	10.06	98	525971	51.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.66%	
62) 4-Bromofluorobenzene	12.38	95	217313	57.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.72%	
Target Compounds						
27) cis-1,2-Dichloroethene	6.79	96	6264	2.080	ug/l	91
44) Trichloroethene	8.80	130	4701	1.619	ug/l	91
64) Tetrachloroethene	10.60	164	118446	41.289	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092820\
Data File : VN063705.D
Acq On : 28 Sep 2020 15:09
Operator : JC/MD
Sample : L4133-02DL 10X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW1DDL

Quant Time: Sep 29 04:08:22 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

