

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092820\
 Data File : VN063707.D
 Acq On : 28 Sep 2020 15:57
 Operator : JC/MD
 Sample : PB131914TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB131914TB

Quant Time: Sep 29 04:22:32 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	207314	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	381370	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	406474	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	163052	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	178243	50.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.02%	
35) Dibromofluoromethane	7.55	113	128850	48.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.60%	
50) Toluene-d8	10.06	98	506563	51.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.06%	
62) 4-Bromofluorobenzene	12.38	95	197833	54.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.72%	
Target Compounds						
16) Acetone	3.78	43	15566	13.725	ug/l	98
20) Methylene Chloride	4.51	84	13419	4.670	ug/l	92
25) 2-Butanone	6.80	43	11053	6.337	ug/l	100
43) Isopropyl Acetate	8.13	43	71066	9.967	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092820\
Data File : VN063707.D
Acq On : 28 Sep 2020 15:57
Operator : JC/MD
Sample : PB131914TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB131914TB

Quant Time: Sep 29 04:22:32 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

