

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092820\
 Data File : VN063708.D
 Acq On : 28 Sep 2020 16:21
 Operator : JC/MD
 Sample : PB131914ZHE#03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB131914ZHE#03

Quant Time: Sep 29 04:24:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	217675	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	411019	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	430905	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	180489	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	191185	51.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.20%	
35) Dibromofluoromethane	7.55	113	135641	47.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.34%	
50) Toluene-d8	10.06	98	536089	50.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.22%	
62) 4-Bromofluorobenzene	12.38	95	217038	55.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.68%	
Target Compounds						
16) Acetone	3.78	43	15516	13.029	ug/l	97
20) Methylene Chloride	4.51	84	14450	4.789	ug/l	87
25) 2-Butanone	6.79	43	11514	6.287	ug/l	95
43) Isopropyl Acetate	8.13	43	70760	9.208	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092820\
Data File : VN063708.D
Acq On : 28 Sep 2020 16:21
Operator : JC/MD
Sample : PB131914ZHE#03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB131914ZHE#03

Quant Time: Sep 29 04:24:58 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

