

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092822\
 Data File : VN074643.D
 Acq On : 28 Sep 2022 12:32
 Operator : JC\MD
 Sample : VN0928WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0928WBL01

Quant Time: Sep 28 18:17:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N092722W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Sep 28 07:48:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.492	128	105109	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	8.827	114	633945	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.621	117	505669	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.292	65	307422	31.019	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	103.400%
60) 4-Bromofluorobenzene	12.610	95	232697	23.690	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	78.967%
63) Toluene-d8	10.310	98	921921	32.323	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	107.733%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092822\
Data File : VN074643.D
Acq On : 28 Sep 2022 12:32
Operator : JC\MD
Sample : VN0928WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0928WBL01

Quant Time: Sep 28 18:17:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N092722W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Sep 28 07:48:31 2022
Response via : Initial Calibration

