

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092823\
 Data File : VN079368.D
 Acq On : 28 Sep 2023 12:52
 Operator : JC\MD
 Sample : 04620-13
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STONE-OUT

Quant Time: Sep 29 04:34:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 05:05:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	151186	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	292181	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	271083	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	73818	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	121665	59.614	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	119.220%
35) Dibromofluoromethane	8.171	113	94297	48.518	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.040%
50) Toluene-d8	10.571	98	332847	46.641	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.280%
62) 4-Bromofluorobenzene	12.853	95	105215	46.221	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.440%
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	19417	22.402	ug/l	92
20) Methylene Chloride	5.283	84	11501	5.453	ug/l #	76
43) Isopropyl Acetate	8.694	43	123533	25.338	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092823\
Data File : VN079368.D
Acq On : 28 Sep 2023 12:52
Operator : JC\MD
Sample : 04620-13
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STONE-OUT

Quant Time: Sep 29 04:34:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 05:05:24 2023
Response via : Initial Calibration

