

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092823\
 Data File : VN079372.D
 Acq On : 28 Sep 2023 14:27
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Sep 29 04:36:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 05:05:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	115903	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	225013	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	187272	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	46792	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	112846	72.125	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	144.260%#
35) Dibromofluoromethane	8.171	113	83191	55.581	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	111.160%
50) Toluene-d8	10.571	98	238612	43.417	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	86.840%
62) 4-Bromofluorobenzene	12.853	95	69934	39.893	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	79.780%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092823\
Data File : VN079372.D
Acq On : 28 Sep 2023 14:27
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Sep 29 04:36:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 05:05:24 2023
Response via : Initial Calibration

