

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092920\
 Data File : VN063727.D
 Acq On : 29 Sep 2020 15:04
 Operator : JC/MD
 Sample : PB131952TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB131952TB

Quant Time: Sep 30 01:10:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	210208	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	378842	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	389215	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	161227	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	173052	48.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.72%	
35) Dibromofluoromethane	7.55	113	126315	48.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.32%	
50) Toluene-d8	10.06	98	501062	50.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.62%	
62) 4-Bromofluorobenzene	12.38	95	193095	53.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.82%	
Target Compounds						
16) Acetone	3.78	43	18467	16.058	ug/l	99
20) Methylene Chloride	4.50	84	19028	6.531	ug/l	91
25) 2-Butanone	6.80	43	14196	8.026	ug/l #	90
43) Isopropyl Acetate	8.13	43	69135	9.761	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092920\
Data File : VN063727.D
Acq On : 29 Sep 2020 15:04
Operator : JC/MD
Sample : PB131952TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB131952TB

Quant Time: Sep 30 01:10:46 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

