

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092920\
 Data File : VN063733.D
 Acq On : 29 Sep 2020 17:52
 Operator : JC/MD
 Sample : L3867-12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EO-01-092820

Manual Integrations
 APPROVED

MMDadoda
 9/30/2020 1:20:45 PM

Quant Time: Sep 30 01:46:44 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	242019	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	437054	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	452697	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	197381	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	193324	46.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.86%	
35) Dibromofluoromethane	7.55	113	145022	47.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.86%	
50) Toluene-d8	10.06	98	562892	49.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.96%	
62) 4-Bromofluorobenzene	12.38	95	228143	54.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.40%	
Target Compounds						
16) Acetone	3.78	43	22057	16.659	ug/l	96
18) Methyl Acetate	4.29	43	3512m	1.045	ug/l	
20) Methylene Chloride	4.51	84	9305m	2.774	ug/l	
43) Isopropyl Acetate	8.13	43	67069	8.208	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092920\
Data File : VN063733.D
Acq On : 29 Sep 2020 17:52
Operator : JC/MD
Sample : L3867-12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EO-01-092820

Manual Integrations
APPROVED

MMDadoda
9/30/2020 1:20:45 PM

Quant Time: Sep 30 01:46:44 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

