

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092920\
 Data File : VN063734.D
 Acq On : 29 Sep 2020 18:15
 Operator : JC/MD
 Sample : L3869-15
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OR-03-092820

Quant Time: Sep 30 01:49:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	226983	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	408753	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	431192	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	183508	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	185707	48.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.12%	
35) Dibromofluoromethane	7.55	113	137175	48.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.96%	
50) Toluene-d8	10.06	98	535982	50.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.74%	
62) 4-Bromofluorobenzene	12.38	95	216537	55.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.02%	
Target Compounds						
16) Acetone	3.78	43	108389	87.285	ug/l	100
20) Methylene Chloride	4.51	84	9319	2.962	ug/l	90
43) Isopropyl Acetate	8.13	43	20492	2.681	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092920\
Data File : VN063734.D
Acq On : 29 Sep 2020 18:15
Operator : JC/MD
Sample : L3869-15
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OR-03-092820

Quant Time: Sep 30 01:49:46 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

