

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092921\
 Data File : VN068746.D
 Acq On : 29 Sep 2021 18:15
 Operator : JC/MD
 Sample : M3986-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 P287-DITCH

Quant Time: Sep 29 23:42:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092921W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 29 23:31:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	504205	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	868552	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	738502	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	257653	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	335613	47.773	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.540%
35) Dibromofluoromethane	8.021	113	264087	48.937	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.880%
50) Toluene-d8	10.443	98	1083497	54.944	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	109.880%
62) 4-Bromofluorobenzene	12.736	95	282554	41.236	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	82.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092921\
Data File : VN068746.D
Acq On : 29 Sep 2021 18:15
Operator : JC/MD
Sample : M3986-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
P287-DITCH

Quant Time: Sep 29 23:42:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092921W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 29 23:31:35 2021
Response via : Initial Calibration

