

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092921\
 Data File : VN068750.D
 Acq On : 29 Sep 2021 19:56
 Operator : JC/MD
 Sample : M3915-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT101D1-20210921

Quant Time: Sep 29 23:43:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092921W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 29 23:31:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	507600	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	863830	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	728696	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	256419	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	342379	48.410	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	96.820%		
35) Dibromofluoromethane	8.021	113	270238	50.350	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	100.700%		
50) Toluene-d8	10.446	98	1069824	54.548	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	109.100%		
62) 4-Bromofluorobenzene	12.734	95	285403	41.879	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	83.760%		
Target Compounds						Qvalue
2) Dichlorodifluoromethane	2.073	85	5060	0.910	ug/l	92
9) 1,1,2-Trichlorotrifluo...	4.218	101	60579	10.883	ug/l	94
12) 1,1-Dichloroethene	4.186	96	21731	4.556	ug/l	93
27) cis-1,2-Dichloroethene	7.327	96	8746	1.269	ug/l	68
30) Chloroform	7.820	83	9470	0.756	ug/l	92
38) Carbon Tetrachloride	8.206	117	8795	0.911	ug/l	93
44) Trichloroethene	9.217	130	1148577	173.942	ug/l	86
55) 1,1,2-Trichloroethane	10.902	97	2718	0.388	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092921\
Data File : VN068750.D
Acq On : 29 Sep 2021 19:56
Operator : JC/MD
Sample : M3915-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT101D1-20210921

Quant Time: Sep 29 23:43:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092921W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 29 23:31:35 2021
Response via : Initial Calibration

