

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092921\
 Data File : VN068751.D
 Acq On : 29 Sep 2021 20:21
 Operator : JC/MD
 Sample : M3915-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT101D-20210921

Quant Time: Sep 29 23:43:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092921W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 29 23:31:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	532150	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	909250	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	724472	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	264548	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	359266	48.454	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	96.900%		
35) Dibromofluoromethane	8.024	113	287079	50.816	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	101.640%		
50) Toluene-d8	10.443	98	1087426	52.675	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	105.360%		
62) 4-Bromofluorobenzene	12.734	95	291170	40.591	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	81.180%		
Target Compounds						Qvalue
2) Dichlorodifluoromethane	2.075	85	6273	1.076	ug/l	94
9) 1,1,2-Trichlorotrifluo...	4.216	101	54156	9.280	ug/l	93
12) 1,1-Dichloroethene	4.189	96	9648	1.930	ug/l #	75
27) cis-1,2-Dichloroethene	7.329	96	18051	2.497	ug/l	75
30) Chloroform	7.817	83	4654	0.354	ug/l	88
44) Trichloroethene	9.217	130	608365	87.529	ug/l	85
64) Tetrachloroethene	10.980	164	4116	0.761	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092921\
Data File : VN068751.D
Acq On : 29 Sep 2021 20:21
Operator : JC/MD
Sample : M3915-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT101D-20210921

Quant Time: Sep 29 23:43:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092921W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 29 23:31:35 2021
Response via : Initial Calibration

