

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092921\
 Data File : VN068753.D
 Acq On : 29 Sep 2021 21:11
 Operator : JC/MD
 Sample : M3915-04DL 20X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE137-20210921DL

Quant Time: Sep 29 23:44:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092921W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 29 23:31:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	485705	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.968	114	833988	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	727800	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	251781	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	335605	49.591	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.180%
35) Dibromofluoromethane	8.024	113	263609	50.872	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.740%
50) Toluene-d8	10.443	98	1045726	55.227	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	110.460%
62) 4-Bromofluorobenzene	12.734	95	281388	42.767	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	85.540%
Target Compounds						
44) Trichloroethene	9.218	130	425452	66.736	ug/l	Qvalue 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092921\
Data File : VN068753.D
Acq On : 29 Sep 2021 21:11
Operator : JC/MD
Sample : M3915-04DL 20X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE137-20210921DL

Quant Time: Sep 29 23:44:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092921W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 29 23:31:35 2021
Response via : Initial Calibration

