

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092922\
 Data File : VN074651.D
 Acq On : 29 Sep 2022 09:28
 Operator : JC\MD
 Sample : VSTDICC001
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC001

Quant Time: Sep 30 07:06:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092922W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 30 07:05:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.934	168	431725	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.822	114	776986	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.616	117	788865	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.545	152	364433	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.292	65	12824	1.185	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	2.380%#
35) Dibromofluoromethane	7.875	113	8287	1.008	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	2.020%#
50) Toluene-d8	10.310	98	34787	1.006	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	2.020%#
62) 4-Bromofluorobenzene	12.610	95	10333	0.942	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	1.880%#
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.999	85	8959	0.991	ug/l	84
3) Chloromethane	2.216	50	14614	1.105	ug/l	91
4) Vinyl Chloride	2.352	62	20638	1.075	ug/l #	85
5) Bromomethane	2.734	94	16647	1.255	ug/l	92
7) Trichlorofluoromethane	3.246	101	15870	1.189	ug/l #	78
8) Diethyl Ether	3.675	74	8067	1.274	ug/l	77
9) 1,1,2-Trichlorotrifluo...	4.051	101	9133	1.148	ug/l #	45
10) Methyl Iodide	4.257	142	8842	0.935	ug/l #	88
11) Tert butyl alcohol	5.175	59	11802	4.084	ug/l	95
12) 1,1-Dichloroethene	4.016	96	9778	1.213	ug/l	85
13) Acrolein	3.899	56	11373	9.108	ug/l #	82
14) Allyl chloride	4.651	41	20213	1.161	ug/l	93
15) Acrylonitrile	5.357	53	35380	5.189	ug/l	95
16) Acetone	4.128	43	37352	6.396	ug/l	98
17) Carbon Disulfide	4.363	76	26593	1.130	ug/l	98
18) Methyl Acetate	4.675	43	23929	1.197	ug/l #	76
19) Methyl tert-butyl Ether	5.416	73	33596	1.092	ug/l #	90
20) Methylene Chloride	4.887	84	7613	0.687	ug/l #	69
21) trans-1,2-Dichloroethene	5.404	96	10398	1.180	ug/l #	65
22) Diisopropyl ether	6.304	45	36823	1.039	ug/l #	94
23) Vinyl Acetate	6.240	43	159687	5.128	ug/l #	92
24) 1,1-Dichloroethane	6.192	63	18835	1.042	ug/l #	80
25) 2-Butanone	7.163	43	53221	5.344	ug/l	90
26) 2,2-Dichloropropane	7.157	77	9067	0.590	ug/l #	53
27) cis-1,2-Dichloroethene	7.145	96	12955	1.202	ug/l	95
28) Bromochloromethane	7.492	49	4806	0.585	ug/l #	79
29) Tetrahydrofuran	7.522	42	33977	5.171	ug/l	90
30) Chloroform	7.651	83	15796	0.948	ug/l	84
31) Cyclohexane	7.928	56	27647	1.573	ug/l #	63
32) 1,1,1-Trichloroethane	7.851	97	17271	1.181	ug/l	92
36) 1,1-Dichloropropene	8.063	75	15284	1.160	ug/l	94
37) Ethyl Acetate	7.234	43	6197	0.317	ug/l #	30
38) Carbon Tetrachloride	8.039	117	12391	1.027	ug/l #	83
39) Methylcyclohexane	9.322	83	16058	0.977	ug/l #	72
40) Benzene	8.310	78	40912	1.008	ug/l	93
41) Methacrylonitrile	7.469	41	8610	1.009	ug/l #	90

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092922\
 Data File : VN074651.D
 Acq On : 29 Sep 2022 09:28
 Operator : JC\MD
 Sample : VSTDICC001
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC001

Quant Time: Sep 30 07:06:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092922W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 30 07:05:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
42) 1,2-Dichloroethane	8.375	62	15538	1.139	ug/l	98
43) Isopropyl Acetate	8.416	43	34037	1.092	ug/l #	91
44) Trichloroethene	9.075	130	7929	0.938	ug/l	82
45) 1,2-Dichloropropane	9.345	63	12168	1.101	ug/l	96
46) Dibromomethane	9.433	93	7734	1.110	ug/l #	45
47) Bromodichloromethane	9.622	83	14585	1.071	ug/l #	91
48) Methyl methacrylate	9.422	41	12876	1.003	ug/l	93
49) 1,4-Dioxane	9.428	88	5234	21.369	ug/l #	91
51) 4-Methyl-2-Pentanone	10.198	43	97830	4.962	ug/l	91
52) Toluene	10.369	92	24229	0.988	ug/l	100
53) t-1,3-Dichloropropene	10.598	75	14287	0.926	ug/l	95
54) cis-1,3-Dichloropropene	10.057	75	17791	1.046	ug/l	98
55) 1,1,2-Trichloroethane	10.775	97	11553	1.170	ug/l #	80
56) Ethyl methacrylate	10.633	69	15420	0.905	ug/l #	71
57) 1,3-Dichloropropane	10.922	76	17270	0.988	ug/l	95
58) 2-Chloroethyl Vinyl ether	9.910	63	38410	5.521	ug/l #	86
59) 2-Hexanone	10.963	43	75194	5.046	ug/l	87
60) Dibromochloromethane	11.110	129	9525	1.002	ug/l	92
61) 1,2-Dibromoethane	11.216	107	9121	0.942	ug/l	97
64) Tetrachloroethene	10.851	164	7799	1.169	ug/l #	75
65) Chlorobenzene	11.645	112	24835	1.052	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.722	131	7521	0.893	ug/l #	52
67) Ethyl Benzene	11.716	91	42204	0.956	ug/l	94
68) m/p-Xylenes	11.822	106	33395	1.997	ug/l	96
69) o-Xylene	12.151	106	16321	0.971	ug/l	96
70) Styrene	12.174	104	24279	0.906	ug/l	92
71) Bromoform	12.333	173	7400	1.096	ug/l #	85
73) Isopropylbenzene	12.457	105	36690	1.075	ug/l	95
74) N-amyl acetate	12.269	43	26501	1.267	ug/l #	86
75) 1,1,2,2-Tetrachloroethane	12.710	83	16327	1.257	ug/l	98
76) 1,2,3-Trichloropropane	12.757	75	20205	2.134	ug/l #	1
77) Bromobenzene	12.739	156	9648	1.319	ug/l	95
78) n-propylbenzene	12.798	91	40889	1.038	ug/l	95
79) 2-Chlorotoluene	12.880	91	27346	1.156	ug/l	96
80) 1,3,5-Trimethylbenzene	12.933	105	26524	0.974	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.510	75	5568	1.277	ug/l	94
82) 4-Chlorotoluene	12.980	91	23500	1.067	ug/l	94
83) tert-Butylbenzene	13.198	119	26658	1.096	ug/l	97
84) 1,2,4-Trimethylbenzene	13.245	105	26934	0.990	ug/l	95
85) sec-Butylbenzene	13.374	105	34821	1.018	ug/l	99
86) p-Isopropyltoluene	13.492	119	27514	1.016	ug/l	98
87) 1,3-Dichlorobenzene	13.486	146	14657	1.105	ug/l	93
88) 1,4-Dichlorobenzene	13.568	146	15931	1.175	ug/l	94
89) n-Butylbenzene	13.816	91	24260	1.015	ug/l	93
90) Hexachloroethane	14.086	117	6704	1.150	ug/l	71
91) 1,2-Dichlorobenzene	13.863	146	16119	1.164	ug/l	92
92) 1,2-Dibromo-3-Chloropr...	14.480	75	3346	1.278	ug/l	89
93) 1,2,4-Trichlorobenzene	15.133	180	6501	1.051	ug/l	96
94) Hexachlorobutadiene	15.233	225	4167	1.376	ug/l	79
95) Naphthalene	15.362	128	22336	0.923	ug/l	97
96) 1,2,3-Trichlorobenzene	15.551	180	3987	0.647	ug/l #	46


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092922\
Data File : VN074651.D
Acq On : 29 Sep 2022 09:28
Operator : JC\MD
Sample : VSTDICC001
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Quant Time: Sep 30 07:06:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092922W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 30 07:05:22 2022
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)





(#)=qualifier out of range (m)=manual integration (+)=signals summed						



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.934 min Scan# 1164
Delta R.T. 0.001 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:168 Resp: 431725
Ion Ratio Lower Upper
168 100
99 57.3 42.3 63.5



#2
Dichlorodifluoromethane
Concen: 0.991 ug/l
RT: 1.999 min Scan# 155
Delta R.T. 0.000 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28

Tgt Ion: 85 Resp: 8959
Ion Ratio Lower Upper
85 100
87 22.7 15.7 47.1









#3
Chloromethane
Concen: 1.105 ug/l
RT: 2.216 min Scan# 14
Delta R.T. 0.000 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001


Tgt Ion: 50 Resp: 14614
Ion Ratio Lower Upper
50 100
52 27.0 25.7 38.5



#4
Vinyl Chloride
Concen: 1.075 ug/l
RT: 2.352 min Scan# 215
Delta R.T. 0.000 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28

Tgt Ion: 62 Resp: 20638
Ion Ratio Lower Upper
62 100
64 39.2 25.0 37.4#









#5
Bromomethane
Concen: 1.255 ug/l
RT: 2.734 min Scan# 21
Delta R.T. 0.018 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001


Tgt Ion: 94 Resp: 16647
Ion Ratio Lower Upper
94 100
96 102.8 76.0 114.0



#7
Trichlorofluoromethane
Concen: 1.189 ug/l
RT: 3.246 min Scan# 367
Delta R.T. 0.006 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28

Tgt Ion: 101 Resp: 15870
Ion Ratio Lower Upper
101 100
103 81.5 51.5 77.3#









#8
Diethyl Ether
Concen: 1.274 ug/l
RT: 3.675 min Scan# 440
Delta R.T. 0.000 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 74 Resp: 8067
Ion Ratio Lower Upper
74 100
45 107.1 42.9 128.5



#9
1,1,2-Trichlorotrifluoroethane
Concen: 1.148 ug/l
RT: 4.051 min Scan# 504
Delta R.T. 0.006 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28

Tgt Ion: 101 Resp: 9133
Ion Ratio Lower Upper
101 100
85 0.0 34.7 52.1#
151 37.0 63.0 94.4#









#10
Methyl Iodide
Concen: 0.935 ug/l
RT: 4.257 min Scan# 51
Delta R.T. 0.006 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:142 Resp: 8842
Ion Ratio Lower Upper
142 100
127 44.2 31.3 46.9
141 5.2 11.2 16.8#



#11
Tert butyl alcohol
Concen: 4.084 ug/l
RT: 5.175 min Scan# 695
Delta R.T. 0.006 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28

Tgt Ion: 59 Resp: 11802
Ion Ratio Lower Upper
59 100
57 12.1 8.3 12.5









#12
1,1-Dichloroethene
Concen: 1.213 ug/l
RT: 4.016 min Scan# 498
Delta R.T. 0.000 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001


Tgt Ion: 96 Resp: 9778
Ion Ratio Lower Upper
96 100
61 136.6 127.6 191.4
98 55.6 51.2 76.8



#13
Acrolein
Concen: 9.108 ug/l
RT: 3.899 min Scan# 478
Delta R.T. 0.012 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28

Tgt Ion: 56 Resp: 11373
Ion Ratio Lower Upper
56 100
55 83.5 54.8 82.2#









#14
 Allyl chloride
 Concen: 1.161 ug/l
 RT: 4.651 min Scan# 60
 Delta R.T. -0.006 min
 Lab File: VN074651.D
 Acq: 29 Sep 2022 09:28




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC001


Tgt Ion: 41	Resp: 20213
Ion Ratio	Lower Upper
41 100	
39 61.4	53.2 79.8
76 32.2	29.9 44.9



#15
 Acrylonitrile
 Concen: 5.189 ug/l
 RT: 5.357 min Scan# 726
 Delta R.T. -0.006 min
 Lab File: VN074651.D
 Acq: 29 Sep 2022 09:28

Tgt Ion: 53	Resp: 35380
Ion Ratio	Lower Upper
53 100	
52 89.0	66.4 99.6
51 36.3	28.5 42.7









#16
Acetone
Concen: 6.396 ug/l
RT: 4.128 min Scan# 51
Delta R.T. 0.006 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 43 Resp: 37352
Ion Ratio Lower Upper
43 100
58 33.8 26.0 39.0



#17
Carbon Disulfide
Concen: 1.130 ug/l
RT: 4.363 min Scan# 557
Delta R.T. 0.006 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28

Tgt Ion: 76 Resp: 26593
Ion Ratio Lower Upper
76 100
78 9.6 7.2 10.8









#18
Methyl Acetate
Concen: 1.197 ug/l
RT: 4.675 min Scan# 61
Delta R.T. 0.012 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 43 Resp: 23929
Ion Ratio Lower Upper
43 100
74 14.7 21.6 32.4#



#19
Methyl tert-butyl Ether
Concen: 1.092 ug/l
RT: 5.416 min Scan# 736
Delta R.T. 0.000 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28

Tgt Ion: 73 Resp: 33596
Ion Ratio Lower Upper
73 100
57 26.9 17.6 26.4#









#20
Methylene Chloride
Concen: 0.687 ug/l
RT: 4.887 min Scan# 64
Delta R.T. -0.012 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001


Tgt Ion:	84	Resp:	7613
Ion	Ratio	Lower	Upper
84	100		
49	77.3	88.3	132.5#
51	18.4	28.2	42.2#
86	39.8	50.6	76.0#



#21
trans-1,2-Dichloroethene
Concen: 1.180 ug/l
RT: 5.404 min Scan# 734
Delta R.T. 0.000 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28

Tgt Ion:	96	Resp:	10398
Ion	Ratio	Lower	Upper
96	100		
61	83.6	105.8	158.8#
98	46.1	50.3	75.5#









#22
Diisopropyl ether
Concen: 1.039 ug/l
RT: 6.304 min Scan# 81
Delta R.T. 0.000 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001


Tgt Ion:	45	Resp:	36823
Ion	Ratio	Lower	Upper
45	100		
43	61.5	46.6	70.0
87	24.2	24.3	36.5#
59	10.6	9.6	14.4



#23
Vinyl Acetate
Concen: 5.128 ug/l
RT: 6.240 min Scan# 876
Delta R.T. -0.006 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28

Tgt Ion:	43	Resp:	159687
Ion	Ratio	Lower	Upper
43	100		
86	9.2	10.0	15.0#









#24
1,1-Dichloroethane
Concen: 1.042 ug/l
RT: 6.192 min Scan# 80
Delta R.T. -0.006 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 63 Resp: 18835
Ion Ratio Lower Upper
63 100
98 17.4 3.5 10.4#
100 3.0 2.3 6.8



#25
2-Butanone
Concen: 5.344 ug/l
RT: 7.163 min Scan# 1033
Delta R.T. 0.000 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28

Tgt Ion: 43 Resp: 53221
Ion Ratio Lower Upper
43 100
72 24.1 23.7 35.5









#26
2,2-Dichloropropane
Concen: 0.590 ug/l
RT: 7.157 min Scan# 1032
Delta R.T. 0.006 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 77 Resp: 9067
Ion Ratio Lower Upper
77 100
97 0.0 11.4 34.2#



#27
cis-1,2-Dichloroethene
Concen: 1.202 ug/l
RT: 7.145 min Scan# 1030
Delta R.T. -0.006 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28

Tgt Ion: 96 Resp: 12955
Ion Ratio Lower Upper
96 100
61 135.2 0.0 280.6
98 59.5 0.0 128.4









#28
Bromochloromethane
Concen: 0.585 ug/l
RT: 7.492 min Scan# 1094
Delta R.T. 0.000 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001


Tgt Ion: 49 Resp: 4806
Ion Ratio Lower Upper
49 100
129 0.0 0.0 4.6
130 107.0 69.7 104.5#



#29
Tetrahydrofuran
Concen: 5.171 ug/l
RT: 7.522 min Scan# 1094
Delta R.T. 0.006 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28

Tgt Ion: 42 Resp: 33977
Ion Ratio Lower Upper
42 100
72 40.0 39.7 59.5
71 42.6 36.6 55.0









#30
Chloroform
Concen: 0.948 ug/l
RT: 7.651 min Scan# 1116
Delta R.T. -0.006 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 83 Resp: 15796
Ion Ratio Lower Upper
83 100
85 77.2 51.8 77.8



#31
Cyclohexane
Concen: 1.573 ug/l
RT: 7.928 min Scan# 1163
Delta R.T. -0.006 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28

Tgt Ion: 56 Resp: 27647
Ion Ratio Lower Upper
56 100
69 70.8 27.8 41.8#
84 68.6 76.9 115.3#





#32

1,1,1-Trichloroethane

Concen: 1.181 ug/l

RT: 7.851 min Scan# 1150

Delta R.T. 0.000 min

Lab File: VN074651.D

Acq: 29 Sep 2022 09:28

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001





Tgt Ion: 97 Resp: 17271

Ion Ratio Lower Upper

97 100

99 53.1 50.6 76.0

61 44.0 36.0 54.0



#33

1,2-Dichloroethane-d4

Concen: 1.185 ug/l

RT: 8.292 min Scan# 1225

Delta R.T. 0.006 min

Lab File: VN074651.D




Acq: 29 Sep 2022 09:28


Tgt Ion: 65 Resp: 12824

Ion Ratio Lower Upper

65 100

67 50.8 0.0 104.8





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.822 min Scan# 1180

Delta R.T. -0.006 min

Lab File: VN074651.D

Acq: 29 Sep 2022 09:28

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001





Tgt Ion:114 Resp: 776986

Ion Ratio Lower Upper

114 100

63 22.2 0.0 40.0

88 13.8 0.0 32.6



#35

Dibromofluoromethane

Concen: 1.008 ug/l

RT: 7.875 min Scan# 1154

Delta R.T. 0.006 min

Lab File: VN074651.D

Acq: 29 Sep 2022 09:28




Tgt Ion:113 Resp: 8287


Ion Ratio Lower Upper

113 100

111 116.8 81.4 122.0

192 12.9 17.0 25.6#









#36
1,1-Dichloropropene
Concen: 1.160 ug/l
RT: 8.063 min Scan# 11
Delta R.T. -0.006 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 75	Resp: 15284
Ion Ratio	Lower Upper
75 100	
110 34.8	18.3 54.8
77 26.5	25.1 37.7



#37
Ethyl Acetate
Concen: 0.317 ug/l
RT: 7.234 min Scan# 1045
Delta R.T. -0.012 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28

Tgt Ion: 43	Resp: 6197
Ion Ratio	Lower Upper
43 100	
61 42.9	11.8 17.6#
70 39.7	9.3 13.9#









#38
Carbon Tetrachloride
Concen: 1.027 ug/l
RT: 8.039 min Scan# 1117
Delta R.T. -0.018 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001


Tgt Ion:117 Resp: 12391
Ion Ratio Lower Upper
117 100
119 74.9 75.8 113.8#
121 26.7 24.6 37.0



#39
Methylcyclohexane
Concen: 0.977 ug/l
RT: 9.322 min Scan# 1400
Delta R.T. 0.006 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28

Tgt Ion: 83 Resp: 16058
Ion Ratio Lower Upper
83 100
55 44.3 57.8 86.8#
98 35.6 39.4 59.2#









#40
Benzene
Concen: 1.008 ug/l
RT: 8.310 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 78 Resp: 40912
Ion Ratio Lower Upper
78 100
77 20.8 19.4 29.0



#41
Methacrylonitrile
Concen: 1.009 ug/l
RT: 7.469 min Scan# 1085
Delta R.T. -0.006 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28

Tgt Ion: 41 Resp: 8610
Ion Ratio Lower Upper
41 100
39 79.3 48.6 73.0#
67 91.8 73.4 110.2
52 40.3 27.7 41.5









#42
1,2-Dichloroethane
Concen: 1.139 ug/l
RT: 8.375 min Scan# 1104
Delta R.T. -0.006 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 62 Resp: 15538
Ion Ratio Lower Upper
62 100
98 8.8 0.0 19.4



#43
Isopropyl Acetate
Concen: 1.092 ug/l
RT: 8.416 min Scan# 1246
Delta R.T. 0.000 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28

Tgt Ion: 43 Resp: 34037
Ion Ratio Lower Upper
43 100
61 23.7 21.9 32.9
87 9.8 12.0 18.0#





#44

Trichloroethene

Concen: 0.938 ug/l

RT: 9.075 min Scan# 11

Delta R.T. 0.000 min

Lab File: VN074651.D

Acq: 29 Sep 2022 09:28

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC001

Tgt Ion:130 Resp: 7929

Ion Ratio Lower Upper

130 100

95 110.5 0.0 186.8



#45

1,2-Dichloropropane

Concen: 1.101 ug/l

RT: 9.345 min Scan# 1404

Delta R.T. -0.006 min

Lab File: VN074651.D




Acq: 29 Sep 2022 09:28


Tgt Ion: 63 Resp: 12168

Ion Ratio Lower Upper

63 100

65 33.4 24.9 37.3









#46
Dibromomethane
Concen: 1.110 ug/l
RT: 9.433 min Scan# 1419
Delta R.T. -0.006 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 93 Resp: 7734
Ion Ratio Lower Upper
93 100
95 0.0 66.9 100.3#
174 81.9 86.6 130.0#



#47
Bromodichloromethane
Concen: 1.071 ug/l
RT: 9.622 min Scan# 1451
Delta R.T. -0.006 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28

Tgt Ion: 83 Resp: 14585
Ion Ratio Lower Upper
83 100
85 68.6 50.5 75.7
127 0.0 6.8 10.2#





#48

Methyl methacrylate

Concen: 1.003 ug/l

RT: 9.422 min Scan# 1417

Delta R.T. -0.006 min

Lab File: VN074651.D

Acq: 29 Sep 2022 09:28

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC001





Tgt Ion: 41 Resp: 12876

Ion Ratio Lower Upper

41 100

69 85.8 74.7 112.1

39 61.9 45.6 68.4



#49

1,4-Dioxane

Concen: 21.369 ug/l

RT: 9.428 min Scan# 1418

Delta R.T. -0.006 min

Lab File: VN074651.D

Acq: 29 Sep 2022 09:28




Tgt Ion: 88 Resp: 5234


Ion Ratio Lower Upper

88 100

43 40.7 20.2 30.4#

58 67.4 54.2 81.4









#50
Toluene-d8
Concen: 1.006 ug/l
RT: 10.310 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001


Tgt Ion: 98 Resp: 34787
Ion Ratio Lower Upper
98 100
100 58.1 52.6 78.8



#51
4-Methyl-2-Pentanone
Concen: 4.962 ug/l
RT: 10.198 min Scan# 1549
Delta R.T. -0.006 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28

Tgt Ion: 43 Resp: 97830
Ion Ratio Lower Upper
43 100
58 36.7 34.1 51.1









#52
Toluene
Concen: 0.988 ug/l
RT: 10.369 min Scan# 11
Delta R.T. -0.006 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 92 Resp: 24229
Ion Ratio Lower Upper
92 100
91 172.5 138.6 207.8



#53
t-1,3-Dichloropropene
Concen: 0.926 ug/l
RT: 10.598 min Scan# 1617
Delta R.T. 0.006 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28

Tgt Ion: 75 Resp: 14287
Ion Ratio Lower Upper
75 100
77 34.7 25.6 38.4









#54
cis-1,3-Dichloropropene
Concen: 1.046 ug/l
RT: 10.057 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:	75	Resp:	17791
Ion	Ratio	Lower	Upper
75	100		
77	29.9	25.2	37.8
39	45.6	35.7	53.5



#55
1,1,2-Trichloroethane
Concen: 1.170 ug/l
RT: 10.775 min Scan# 1647
Delta R.T. 0.000 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28

Tgt Ion:	97	Resp:	11553
Ion	Ratio	Lower	Upper
97	100		
83	108.3	66.2	99.4#
85	54.2	43.9	65.9
99	43.2	50.6	76.0#









#56
Ethyl methacrylate
Concen: 0.905 ug/l
RT: 10.633 min Scan# 1623
Delta R.T. -0.006 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 69 Resp: 15420
Ion Ratio Lower Upper
69 100
41 80.7 47.8 71.6#
39 51.4 27.3 40.9#



#57
1,3-Dichloropropane
Concen: 0.988 ug/l
RT: 10.922 min Scan# 1672
Delta R.T. 0.006 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28

Tgt Ion: 76 Resp: 17270
Ion Ratio Lower Upper
76 100
78 29.9 26.2 39.2









#58
2-Chloroethyl Vinyl ether
Concen: 5.521 ug/l
RT: 9.910 min Scan# 11
Delta R.T. -0.006 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001


Tgt Ion: 63 Resp: 38410
Ion Ratio Lower Upper
63 100
106 23.3 24.6 36.8#



#59
2-Hexanone
Concen: 5.046 ug/l
RT: 10.963 min Scan# 1679
Delta R.T. 0.000 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28

Tgt Ion: 43 Resp: 75194
Ion Ratio Lower Upper
43 100
58 49.7 29.7 89.1









#60
Dibromochloromethane
Concen: 1.002 ug/l
RT: 11.110 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:129 Resp: 9525
Ion Ratio Lower Upper
129 100
127 71.1 38.8 116.4



#61
1,2-Dibromoethane
Concen: 0.942 ug/l
RT: 11.216 min Scan# 1722
Delta R.T. -0.006 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28

Tgt Ion:107 Resp: 9121
Ion Ratio Lower Upper
107 100
109 97.1 75.0 112.6









#62
4-Bromofluorobenzene
Concen: 0.942 ug/l
RT: 12.610 min Scan# 1959
Delta R.T. 0.006 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001


Tgt Ion: 95 Resp: 10333
Ion Ratio Lower Upper
95 100
174 83.9 0.0 167.6
176 82.8 0.0 163.4



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.616 min Scan# 1790
Delta R.T. 0.000 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28

Tgt Ion: 117 Resp: 788865
Ion Ratio Lower Upper
117 100
82 59.3 44.2 66.2
119 33.5 25.4 38.0









#64
Tetrachloroethene
Concen: 1.169 ug/l
RT: 10.851 min Scan# 1795
Delta R.T. 0.000 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:164 Resp: 7799
Ion Ratio Lower Upper
164 100
166 85.2 104.2 156.2#
129 76.2 73.9 110.9
131 74.3 69.4 104.2



#65
Chlorobenzene
Concen: 1.052 ug/l
RT: 11.645 min Scan# 1795
Delta R.T. 0.000 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28

Tgt Ion:112 Resp: 24835
Ion Ratio Lower Upper
112 100
114 32.4 25.4 38.2









#66
1,1,1,2-Tetrachloroethane
Concen: 0.893 ug/l
RT: 11.722 min Scan# 11
Delta R.T. 0.006 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001


Tgt Ion:131 Resp: 7521
Ion Ratio Lower Upper
131 100
133 118.5 47.8 143.4
119 0.0 33.4 100.2#



#67
Ethyl Benzene
Concen: 0.956 ug/l
RT: 11.716 min Scan# 1807
Delta R.T. -0.006 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28

Tgt Ion: 91 Resp: 42204
Ion Ratio Lower Upper
91 100
106 34.7 24.9 37.3









#68
m/p-Xylenes
Concen: 1.997 ug/l
RT: 11.822 min Scan# 1181
Delta R.T. -0.006 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:106 Resp: 33395
Ion Ratio Lower Upper
106 100
91 204.1 158.9 238.3



#69
o-Xylene
Concen: 0.971 ug/l
RT: 12.151 min Scan# 1881
Delta R.T. -0.006 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28

Tgt Ion:106 Resp: 16321
Ion Ratio Lower Upper
106 100
91 218.6 106.0 318.0









#70
Styrene
Concen: 0.906 ug/l
RT: 12.174 min Scan# 11
Delta R.T. 0.006 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:104 Resp: 24279
Ion Ratio Lower Upper
104 100
78 40.8 39.5 59.3
103 51.6 44.0 66.0



#71
Bromoform
Concen: 1.096 ug/l
RT: 12.333 min Scan# 1912
Delta R.T. 0.000 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28

Tgt Ion:173 Resp: 7400
Ion Ratio Lower Upper
173 100
175 38.7 24.6 73.6
254 0.0 0.1 0.1#









#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.545 min Scan# 2118
Delta R.T. -0.006 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001


Tgt Ion:152 Resp: 364433
Ion Ratio Lower Upper
152 100
115 63.1 28.9 86.7
150 163.6 0.0 356.2



#73
Isopropylbenzene
Concen: 1.075 ug/l
RT: 12.457 min Scan# 1933
Delta R.T. 0.000 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28

Tgt Ion:105 Resp: 36690
Ion Ratio Lower Upper
105 100
120 29.6 13.5 40.4









#74
N-amy1 acetate
Concen: 1.267 ug/l
RT: 12.269 min Scan# 1901
Delta R.T. 0.000 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Tgt Ion: 43 Resp: 26501




Ion	Ratio	Lower	Upper
43	100		
70	35.1	39.0	58.4#
55	24.5	21.1	31.7
61	19.4	21.1	31.7#




#75
1,1,2,2-Tetrachloroethane
Concen: 1.257 ug/l
RT: 12.710 min Scan# 1976
Delta R.T. 0.006 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28

Tgt Ion: 83 Resp: 16327

Ion	Ratio	Lower	Upper
83	100		
131	12.1	5.3	15.9
85	66.2	32.3	96.9









#76
1,2,3-Trichloropropane
Concen: 2.134 ug/l
RT: 12.757 min Scan# 1848
Delta R.T. 0.000 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 75 Resp: 20205
Ion Ratio Lower Upper
75 100
77 0.0 98.5 295.5#



#77
Bromobenzene
Concen: 1.319 ug/l
RT: 12.739 min Scan# 1981
Delta R.T. 0.006 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28

Tgt Ion:156 Resp: 9648
Ion Ratio Lower Upper
156 100
77 210.5 100.5 301.5
158 94.0 48.9 146.6









#78
n-propylbenzene
Concen: 1.038 ug/l
RT: 12.798 min Scan# 1855
Delta R.T. 0.000 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 91 Resp: 40889
Ion Ratio Lower Upper
91 100
120 21.4 11.9 35.5



#79
2-Chlorotoluene
Concen: 1.156 ug/l
RT: 12.880 min Scan# 2005
Delta R.T. 0.000 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28

Tgt Ion: 91 Resp: 27346
Ion Ratio Lower Upper
91 100
126 32.4 17.3 51.9









#80
1,3,5-Trimethylbenzene
Concen: 0.974 ug/l
RT: 12.933 min Scan# 20
Delta R.T. 0.000 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:105 Resp: 26524
Ion Ratio Lower Upper
105 100
120 50.0 24.9 74.7



#81
trans-1,4-Dichloro-2-butene
Concen: 1.277 ug/l
RT: 12.510 min Scan# 1942
Delta R.T. 0.006 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28

Tgt Ion: 75 Resp: 5568
Ion Ratio Lower Upper
75 100
53 105.3 78.0 117.0
89 47.5 37.2 55.8









#82
4-Chlorotoluene
Concen: 1.067 ug/l
RT: 12.980 min Scan# 2059
Delta R.T. 0.000 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 91 Resp: 23500
Ion Ratio Lower Upper
91 100
126 31.2 17.2 51.6



#83
tert-Butylbenzene
Concen: 1.096 ug/l
RT: 13.198 min Scan# 2059
Delta R.T. 0.000 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28

Tgt Ion: 119 Resp: 26658
Ion Ratio Lower Upper
119 100
91 62.2 31.6 94.8
134 23.1 13.7 41.1









#84
1,2,4-Trimethylbenzene
Concen: 0.990 ug/l
RT: 13.245 min Scan# 2089
Delta R.T. 0.000 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001


Tgt Ion:105 Resp: 26934
Ion Ratio Lower Upper
105 100
120 49.7 23.3 69.8



#85
sec-Butylbenzene
Concen: 1.018 ug/l
RT: 13.374 min Scan# 2089
Delta R.T. 0.000 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28

Tgt Ion:105 Resp: 34821
Ion Ratio Lower Upper
105 100
134 21.0 10.4 31.1









#86
 p-Isopropyltoluene
 Concen: 1.016 ug/l
 RT: 13.492 min Scan# 2109
 Delta R.T. 0.000 min
 Lab File: VN074651.D
 Acq: 29 Sep 2022 09:28




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC001


Tgt Ion:	119	Resp:	27514
Ion	Ratio	Lower	Upper
119	100		
134	24.4	13.3	39.8
91	22.8	11.3	33.9



#87
 1,3-Dichlorobenzene
 Concen: 1.105 ug/l
 RT: 13.486 min Scan# 2108
 Delta R.T. 0.000 min
 Lab File: VN074651.D
 Acq: 29 Sep 2022 09:28

Tgt Ion:	146	Resp:	14657
Ion	Ratio	Lower	Upper
146	100		
111	43.8	20.0	60.0
148	57.7	32.0	96.0









#88
1,4-Dichlorobenzene
Concen: 1.175 ug/l
RT: 13.568 min Scan# 2164
Delta R.T. 0.000 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:146 Resp: 15931
Ion Ratio Lower Upper
146 100
111 32.9 19.1 57.1
148 66.9 31.4 94.3



#89
n-Butylbenzene
Concen: 1.015 ug/l
RT: 13.816 min Scan# 2164
Delta R.T. 0.000 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28

Tgt Ion: 91 Resp: 24260
Ion Ratio Lower Upper
91 100
92 47.0 26.9 80.7
134 26.9 13.6 40.8









#90
Hexachloroethane
Concen: 1.150 ug/l
RT: 14.086 min Scan# 2110
Delta R.T. 0.000 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:117 Resp: 6704
Ion Ratio Lower Upper
117 100
201 55.8 40.6 122.0



#91
1,2-Dichlorobenzene
Concen: 1.164 ug/l
RT: 13.863 min Scan# 2172
Delta R.T. 0.000 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28

Tgt Ion:146 Resp: 16119
Ion Ratio Lower Upper
146 100
111 36.1 20.9 62.8
148 58.0 32.1 96.3









#92
1,2-Dibromo-3-Chloropropane
Concen: 1.278 ug/l
RT: 14.480 min Scan# 2140
Delta R.T. 0.006 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 75 Resp: 3346
Ion Ratio Lower Upper
75 100
155 80.1 44.9 134.5
157 104.7 59.0 176.8



#93
1,2,4-Trichlorobenzene
Concen: 1.051 ug/l
RT: 15.133 min Scan# 2388
Delta R.T. 0.006 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28

Tgt Ion: 180 Resp: 6501
Ion Ratio Lower Upper
180 100
182 94.1 47.5 142.5
145 37.6 14.5 43.5









#94
Hexachlorobutadiene
Concen: 1.376 ug/l
RT: 15.233 min Scan# 2405
Delta R.T. 0.000 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:	225	Resp:	4167
Ion	Ratio	Lower	Upper
225	100		
223	60.3	32.4	97.2
227	35.4	31.8	95.4



#95
Naphthalene
Concen: 0.923 ug/l
RT: 15.362 min Scan# 2427
Delta R.T. 0.000 min
Lab File: VN074651.D
Acq: 29 Sep 2022 09:28


Tgt Ion:	128	Resp:	22336
Ion	Ratio	Lower	Upper
128	100		
127	12.4	10.3	15.5
129	12.6	8.6	13.0





#96
 1,2,3-Trichlorobenzene
 Concen: 0.647 ug/l
 RT: 15.551 min Scan# 24
 Delta R.T. 0.000 min
 Lab File: VN074651.D
 Acq: 29 Sep 2022 09:28

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC001



Tgt Ion:180 Resp: 3987

Ion	Ratio	Lower	Upper
180	100		
182	152.0	47.8	143.3#
145	53.3	15.5	46.5#

