

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092925\
 Data File : VN087977.D
 Acq On : 29 Sep 2025 13:08
 Operator : JC\MD
 Sample : VN0929WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0929WBL01

Quant Time: Sep 30 03:52:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092325W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 24 05:05:30 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.212	168	158754	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.088	114	351007	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	338452	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	154688	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	161783	60.554	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	121.100%
35) Dibromofluoromethane	8.153	113	131718	51.684	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.360%
50) Toluene-d8	10.547	98	448783	50.076	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.160%
62) 4-Bromofluorobenzene	12.829	95	142869	44.598	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	89.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092925\
Data File : VN087977.D
Acq On : 29 Sep 2025 13:08
Operator : JC\MD
Sample : VN0929WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0929WBL01

Quant Time: Sep 30 03:52:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092325W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 24 05:05:30 2025
Response via : Initial Calibration

