

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092925\
 Data File : VN087981.D
 Acq On : 29 Sep 2025 14:46
 Operator : JC\MD
 Sample : Q3208-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SU-ESM-COMP-02

Quant Time: Sep 30 03:55:15 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092325W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 24 05:05:30 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.212	168	176898	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	386906	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	391953	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	170658	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	163027	54.761	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.520%	
35) Dibromofluoromethane	8.147	113	134087	47.732	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 95.460%	
50) Toluene-d8	10.547	98	478748	48.463	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 96.920%	
62) 4-Bromofluorobenzene	12.829	95	155741	44.106	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 88.220%	
Target Compounds						Qvalue
16) Acetone	4.430	43	78560	64.136	ug/l	93
20) Methylene Chloride	5.271	84	34052	13.734	ug/l	95
43) Isopropyl Acetate	8.682	43	17276	2.474	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092925\
Data File : VN087981.D
Acq On : 29 Sep 2025 14:46
Operator : JC\MD
Sample : Q3208-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SU-ESM-COMP-02

Quant Time: Sep 30 03:55:15 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092325W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 24 05:05:30 2025
Response via : Initial Calibration

