

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN093019\
 Data File : VN058423.D
 Acq On : 30 Sep 2019 10:46
 Operator : JC/SP
 Sample : K5080-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB01-20190924

Quant Time: Oct 01 02:13:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 30 08:07:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	447951	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	722725	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	614050	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	242329	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	319147	48.75	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.50%	
35) Dibromofluoromethane	7.57	113	226032	51.03	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.06%	
50) Toluene-d8	10.08	98	899102	51.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.62%	
62) 4-Bromofluorobenzene	12.41	95	293236	46.74	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.48%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN093019\
Data File : VN058423.D
Acq On : 30 Sep 2019 10:46
Operator : JC/SP
Sample : K5080-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB01-20190924

Quant Time: Oct 01 02:13:13 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 30 08:07:38 2019
Response via : Initial Calibration

