

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN093019\
 Data File : VN058426.D
 Acq On : 30 Sep 2019 11:53
 Operator : JC/SP
 Sample : K5080-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SA-MW132I-20190924

Quant Time: Oct 01 07:59:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 30 08:07:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	444157	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	727733	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	611958	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	236587	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	317473	48.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.82%	
35) Dibromofluoromethane	7.57	113	228708	51.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.56%	
50) Toluene-d8	10.09	98	901333	51.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.18%	
62) 4-Bromofluorobenzene	12.41	95	282685	44.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.48%	
Target Compounds						
6) Chloroethane	2.69	64	35225	11.209	ug/l	97
12) 1,1-Dichloroethene	3.72	96	32676	7.799	ug/l	95
16) Acetone	3.81	43	12219	5.522	ug/l	92
24) 1,1-Dichloroethane	5.83	63	1224030	132.548	ug/l	99
32) 1,1,1-Trichloroethane	7.55	97	56882	7.402	ug/l	99
44) Trichloroethene	8.83	130	1652	0.334	ug/l	90
73) Isopropylbenzene	12.25	105	8205	0.469	ug/l	100
85) sec-Butylbenzene	13.18	105	14684	0.903	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN093019\
Data File : VN058426.D
Acq On : 30 Sep 2019 11:53
Operator : JC/SP
Sample : K5080-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SA-MW132I-20190924

Quant Time: Oct 01 07:59:55 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 30 08:07:38 2019
Response via : Initial Calibration

