

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN093019\  
 Data File : VN058428.D  
 Acq On : 30 Sep 2019 12:37  
 Operator : JC/SP  
 Sample : K5080-03  
 Misc : 5.00mL/MSVOA N/WATER  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 FT-PZ461I-20190924

Quant Time: Oct 01 08:04:22 2019  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_N\METHODS\82N092719W.M  
 Quant Title : SW846 8260  
 QLast Update : Mon Sep 30 08:07:38 2019  
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	402064	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	646887	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	545322	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	211588	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	285747	48.63	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.26%	
35) Dibromofluoromethane	7.57	113	201505	50.82	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.64%	
50) Toluene-d8	10.08	98	801543	51.61	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.22%	
62) 4-Bromofluorobenzene	12.41	95	248927	44.33	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.66%	
Target Compounds						
12) 1,1-Dichloroethene	3.72	96	4280	1.128	ug/l #	89
16) Acetone	3.80	43	12159	6.071	ug/l	97
27) cis-1,2-Dichloroethene	6.81	96	11027	2.237	ug/l	86
44) Trichloroethene	8.82	130	99890	22.753	ug/l	97
64) Tetrachloroethene	10.63	164	2815	0.813	ug/l #	84
95) Naphthalene	15.14	128	12664	1.197	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN093019\  
 Data File : VN058428.D  
 Acq On : 30 Sep 2019 12:37  
 Operator : JC/SP  
 Sample : K5080-03  
 Misc : 5.00mL/MSVOA N/WATER  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 Client Sample Id :  
 FT-PZ4611-20190924

Quant Time: Oct 01 08:04:22 2019  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_N\METHODS\82N092719W.M  
 Quant Title : SW846 8260  
 QLast Update : Mon Sep 30 08:07:38 2019  
 Response via : Initial Calibration

