

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN093019\
 Data File : VN058430.D
 Acq On : 30 Sep 2019 13:21
 Operator : JC/SP
 Sample : K5080-09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SA-MW184I-20190924

Manual Integrations
 APPROVED

MMDadoda
 10/2/2019 9:22:21 AM

Quant Time: Oct 01 08:10:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 30 08:07:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	438583	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	728670	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	608691	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	238220	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	320594	50.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.04%	
35) Dibromofluoromethane	7.57	113	225945	50.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.18%	
50) Toluene-d8	10.09	98	907080	51.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.70%	
62) 4-Bromofluorobenzene	12.41	95	282867	44.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.44%	
Target Compounds						
6) Chloroethane	2.69	64	5697	1.836	ug/l	90
12) 1,1-Dichloroethene	3.72	96	15623	3.776	ug/l	98
16) Acetone	3.80	43	13429	6.146	ug/l	89
24) 1,1-Dichloroethane	5.83	63	300069	32.907	ug/l	99
27) cis-1,2-Dichloroethene	6.81	96	2267m	0.422	ug/l	
32) 1,1,1-Trichloroethane	7.56	97	10367	1.366	ug/l #	52
44) Trichloroethene	8.82	130	3359	0.679	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN093019\
Data File : VN058430.D
Acq On : 30 Sep 2019 13:21
Operator : JC/SP
Sample : K5080-09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SA-MW184I-20190924

Manual Integrations
APPROVED

MMDadoda
10/2/2019 9:22:21 AM

Quant Time: Oct 01 08:10:09 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 30 08:07:38 2019
Response via : Initial Calibration

