

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN093019\
 Data File : VN058433.D
 Acq On : 30 Sep 2019 14:27
 Operator : JC/SP
 Sample : K5080-13
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SA-PZ145I-20190924

Quant Time: Oct 01 04:06:34 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 30 08:07:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	432290	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	705054	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	596595	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	228403	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	316959	50.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.34%	
35) Dibromofluoromethane	7.57	113	223220	51.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.32%	
50) Toluene-d8	10.09	98	890790	52.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.24%	
62) 4-Bromofluorobenzene	12.41	95	279508	45.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.34%	
Target Compounds						
12) 1,1-Dichloroethene	3.72	96	2543	0.624	ug/l #	78
16) Acetone	3.80	43	13303	6.177	ug/l	94
24) 1,1-Dichloroethane	5.82	63	29218	3.251	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN093019\
Data File : VN058433.D
Acq On : 30 Sep 2019 14:27
Operator : JC/SP
Sample : K5080-13
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SA-PZ145I-20190924

Quant Time: Oct 01 04:06:34 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 30 08:07:38 2019
Response via : Initial Calibration

