

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN093019\
 Data File : VN058436.D
 Acq On : 30 Sep 2019 15:33
 Operator : JC/SP
 Sample : K5080-16
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SA-DUP01-20190924

Quant Time: Oct 01 04:19:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 30 08:07:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	413741	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	679633	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	578346	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	224233	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	304155	50.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.62%	
35) Dibromofluoromethane	7.57	113	209945	50.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.80%	
50) Toluene-d8	10.09	98	846257	51.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.72%	
62) 4-Bromofluorobenzene	12.41	95	266881	45.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.46%	
Target Compounds						
16) Acetone	3.81	43	13574	6.586	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN093019\
Data File : VN058436.D
Acq On : 30 Sep 2019 15:33
Operator : JC/SP
Sample : K5080-16
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SA-DUP01-20190924

Quant Time: Oct 01 04:19:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 30 08:07:38 2019
Response via : Initial Calibration

