

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN093019\
 Data File : VN058439.D
 Acq On : 30 Sep 2019 16:39
 Operator : JC/SP
 Sample : K5080-19
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SA-MW186S-20190925

Quant Time: Oct 01 04:34:08 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 30 08:07:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	410913	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	670163	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	564776	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	209870	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	302696	50.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.82%	
35) Dibromofluoromethane	7.57	113	210935	51.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.72%	
50) Toluene-d8	10.08	98	850030	52.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.66%	
62) 4-Bromofluorobenzene	12.41	95	262326	45.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.18%	
Target Compounds						
16) Acetone	3.81	43	14158	6.917	ug/l #	86
30) Chloroform	7.35	83	2547	0.294	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN093019\
Data File : VN058439.D
Acq On : 30 Sep 2019 16:39
Operator : JC/SP
Sample : K5080-19
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SA-MW186S-20190925

Quant Time: Oct 01 04:34:08 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 30 08:07:38 2019
Response via : Initial Calibration

