

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN093019\
 Data File : VN058442.D
 Acq On : 30 Sep 2019 17:46
 Operator : JC/SP
 Sample : K5080-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SA-MW184S-20190924

Quant Time: Oct 01 04:47:40 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 30 08:07:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	421110	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	685125	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	582270	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	222351	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	304373	49.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.92%	
35) Dibromofluoromethane	7.58	113	212139	50.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.04%	
50) Toluene-d8	10.08	98	844654	51.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.70%	
62) 4-Bromofluorobenzene	12.41	95	263086	44.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.46%	
Target Compounds						
16) Acetone	3.81	43	13584	6.475	ug/l	99
30) Chloroform	7.36	83	32504	3.664	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN093019\
Data File : VN058442.D
Acq On : 30 Sep 2019 17:46
Operator : JC/SP
Sample : K5080-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SA-MW184S-20190924

Quant Time: Oct 01 04:47:40 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 30 08:07:38 2019
Response via : Initial Calibration

