

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN093019\
 Data File : VN058444.D
 Acq On : 30 Sep 2019 18:30
 Operator : JC/SP
 Sample : K5080-10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SA-MW185I-20190924

Quant Time: Oct 01 04:56:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 30 08:07:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	429216	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	685787	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	594347	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	228167	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	309066	49.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.54%	
35) Dibromofluoromethane	7.57	113	215585	51.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.58%	
50) Toluene-d8	10.08	98	874329	53.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.20%	
62) 4-Bromofluorobenzene	12.41	95	275920	46.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.70%	
Target Compounds						
16) Acetone	3.80	43	12818	5.995	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN093019\
Data File : VN058444.D
Acq On : 30 Sep 2019 18:30
Operator : JC/SP
Sample : K5080-10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SA-MW185I-20190924

Quant Time: Oct 01 04:56:17 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 30 08:07:38 2019
Response via : Initial Calibration

