

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN093019\
 Data File : VN058448.D
 Acq On : 30 Sep 2019 19:58
 Operator : JC/SP
 Sample : K5079-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-1-COMP

Quant Time: Oct 01 05:21:51 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 30 08:07:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	422435	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	700801	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	609814	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	243412	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	304036	49.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.50%	
35) Dibromofluoromethane	7.58	113	217251	50.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.16%	
50) Toluene-d8	10.09	98	882614	52.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.92%	
62) 4-Bromofluorobenzene	12.41	95	287632	47.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.56%	
Target Compounds						
20) Methylene Chloride	4.53	84	364758	75.109	ug/l	99
43) Isopropyl Acetate	8.15	43	96101	8.995	ug/l	100
84) 1,2,4-Trimethylbenzene	13.05	105	17067	1.131	ug/l	98
95) Naphthalene	15.13	128	19603	1.620	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN093019\
Data File : VN058448.D
Acq On : 30 Sep 2019 19:58
Operator : JC/SP
Sample : K5079-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-1-COMP

Quant Time: Oct 01 05:21:51 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 30 08:07:38 2019
Response via : Initial Calibration

