

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN093020\
 Data File : VN063745.D
 Acq On : 30 Sep 2020 13:27
 Operator : JC/MD
 Sample : VN0930MBL01
 Misc : 5.00µ/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0930MBL01

Quant Time: Oct 01 02:26:23 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	242180	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	435464	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	428666	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	183202	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	197072	49.26	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.52%	
35) Dibromofluoromethane	7.55	113	143233	47.40	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.80%	
50) Toluene-d8	10.06	98	550575	48.07	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.14%	
62) 4-Bromofluorobenzene	12.38	95	212012	49.24	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.48%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN093020\
Data File : VN063745.D
Acq On : 30 Sep 2020 13:27
Operator : JC/MD
Sample : VN0930MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0930MBL01

Quant Time: Oct 01 02:26:23 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

