

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN093020\
 Data File : VN063752.D
 Acq On : 30 Sep 2020 16:20
 Operator : JC/MD
 Sample : L4173-01
 Misc : 2.44uL/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-1-(3FT)

Quant Time: Oct 01 02:50:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	219496	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	405013	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	411587	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	196198	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.98	65	181663	50.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.20%	
35) Dibromofluoromethane	7.55	113	128991	45.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.80%	
50) Toluene-d8	10.06	98	516903	48.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.04%	
62) 4-Bromofluorobenzene	12.38	95	216576	54.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.16%	
Target Compounds						
20) Methylene Chloride	4.50	84	4509	1.858	ug/l #	79
95) Naphthalene	15.11	128	21523	3.879	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN093020\
Data File : VN063752.D
Acq On : 30 Sep 2020 16:20
Operator : JC/MD
Sample : L4173-01
Misc : 2.44g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-1-(3FT)

Quant Time: Oct 01 02:50:14 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

