


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN093020\  
 Data File : VN063736.D  
 Acq On : 30 Sep 2020 8:31  
 Operator : JC/MD  
 Sample : BFB  
 Misc : 5.00mL/MSVOA N/WATER  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA\_N\METHODS\82N093020W.M  
 Title : SW846 8260  
 Last Update : Thu Oct 01 01:46:38 2020



AutoFind: Scans 3307, 3308, 3309; Background Corrected with Scan 3293

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 50          | 95           | 15           | 40           | 23.8      | 18008   | PASS             |
| 75          | 95           | 30           | 60           | 57.3      | 43317   | PASS             |
| 95          | 95           | 100          | 100          | 100.0     | 75600   | PASS             |
| 96          | 95           | 5            | 9            | 7.0       | 5327    | PASS             |
| 173         | 174          | 0.00         | 2            | 0.9       | 487     | PASS             |
| 174         | 95           | 50           | 100          | 74.2      | 56114   | PASS             |
| 175         | 174          | 5            | 9            | 8.0       | 4464    | PASS             |
| 176         | 174          | 95           | 101          | 98.6      | 55317   | PASS             |
| 177         | 176          | 5            | 9            | 5.9       | 3250    | PASS             |