

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093021\
 Data File : VN068759.D
 Acq On : 30 Sep 2021 12:58
 Operator : JC/MD
 Sample : VN0930WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0930WBL01

Quant Time: Sep 30 15:18:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092921W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 29 23:31:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	347168	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.968	114	613112	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	558133	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	193177	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	246725	51.006	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 102.020%	
35) Dibromofluoromethane	8.021	113	200002	52.502	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 105.000%	
50) Toluene-d8	10.443	98	691674	49.688	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 99.380%	
62) 4-Bromofluorobenzene	12.734	95	209919	43.399	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 86.800%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093021\
Data File : VN068759.D
Acq On : 30 Sep 2021 12:58
Operator : JC/MD
Sample : VN0930WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0930WBL01

Quant Time: Sep 30 15:18:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092921W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 29 23:31:35 2021
Response via : Initial Calibration

