

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093021\
 Data File : VN068764.D
 Acq On : 30 Sep 2021 15:49
 Operator : JC/MD
 Sample : M3915-02RE
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT101D-20210921RE

Manual Integrations
 APPROVED

MMDadoda
 10/1/2021 11:02:49 AM

Quant Time: Oct 01 01:55:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092921W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 29 23:31:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	498100	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.968	114	859666	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	682691	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	242401	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	345154	49.733	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	99.460%		
35) Dibromofluoromethane	8.024	113	269906	50.532	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	101.060%		
50) Toluene-d8	10.443	98	1021016	52.311	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	104.620%		
62) 4-Bromofluorobenzene	12.733	95	271604	40.047	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	80.100%		
Target Compounds						Qvalue
2) Dichlorodifluoromethane	2.075	85	6350	1.164	ug/l	93
9) 1,1,2-Trichlorotrifluo...	4.216	101	57372	10.503	ug/l	93
12) 1,1-Dichloroethene	4.186	96	9923	2.120	ug/l	80
27) cis-1,2-Dichloroethene	7.329	96	18058m	2.669	ug/l	
30) Chloroform	7.817	83	4142	0.337	ug/l	85
44) Trichloroethene	9.217	130	608598	92.613	ug/l	87
64) Tetrachloroethene	10.982	164	4012	0.787	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093021\
Data File : VN068764.D
Acq On : 30 Sep 2021 15:49
Operator : JC/MD
Sample : M3915-02RE
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT101D-20210921RE

Manual Integrations
APPROVED

MMDadoda
10/1/2021 11:02:49 AM

Quant Time: Oct 01 01:55:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092921W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 29 23:31:35 2021
Response via : Initial Calibration

