

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093021\
 Data File : VN068766.D
 Acq On : 30 Sep 2021 16:39
 Operator : JC/MD
 Sample : M3915-08RE
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RW3-MW1-20210921RE

Quant Time: Oct 01 01:56:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092921W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 29 23:31:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	521488	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	888243	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	701300	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	257634	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	361092	49.696	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.400%
35) Dibromofluoromethane	8.024	113	279646	50.671	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.340%
50) Toluene-d8	10.443	98	1051071	52.118	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.240%
62) 4-Bromofluorobenzene	12.731	95	278740	39.777	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	79.560%
Target Compounds						
					Qvalue	
44) Trichloroethene	9.217	130	51072	7.522	ug/l	86
64) Tetrachloroethene	10.979	164	2629	0.502	ug/l	# 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093021\
Data File : VN068766.D
Acq On : 30 Sep 2021 16:39
Operator : JC/MD
Sample : M3915-08RE
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RW3-MW1-20210921RE

Quant Time: Oct 01 01:56:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092921W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 29 23:31:35 2021
Response via : Initial Calibration

