

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093021\
 Data File : VN068767.D
 Acq On : 30 Sep 2021 17:04
 Operator : JC/MD
 Sample : M3915-07RE
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FB02-20210921RE

Quant Time: Oct 01 01:56:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092921W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 29 23:31:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	484702	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	829424	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	650355	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	239836	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	335929	49.742	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.480%
35) Dibromofluoromethane	8.027	113	265085	51.439	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.880%
50) Toluene-d8	10.443	98	980775	52.082	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.160%
62) 4-Bromofluorobenzene	12.733	95	264985	40.496	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	81.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093021\
Data File : VN068767.D
Acq On : 30 Sep 2021 17:04
Operator : JC/MD
Sample : M3915-07RE
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FB02-20210921RE

Quant Time: Oct 01 01:56:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092921W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 29 23:31:35 2021
Response via : Initial Calibration

