

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093021\
 Data File : VN068772.D
 Acq On : 30 Sep 2021 19:11
 Operator : JC/MD
 Sample : M3915-12
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW17811-20210921

Quant Time: Oct 01 04:49:20 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092921W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 29 23:31:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	482497	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	827416	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	697552	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	235469	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	331894	49.369	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	98.740%		
35) Dibromofluoromethane	8.021	113	257875	50.161	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	100.320%		
50) Toluene-d8	10.443	98	1014503	54.003	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	108.000%		
62) 4-Bromofluorobenzene	12.733	95	294985	45.190	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	90.380%		
Target Compounds						Qvalue
12) 1,1-Dichloroethene	4.186	96	12839	2.832	ug/l	85
27) cis-1,2-Dichloroethene	7.326	96	245593	37.474	ug/l	75
44) Trichloroethene	9.217	130	2969386	469.478	ug/l	87
45) 1,2-Dichloropropane	9.491	63	140470	22.978	ug/l	98
64) Tetrachloroethene	10.979	164	137443	26.396	ug/l	95
76) 1,2,3-Trichloropropane	12.884	75	17060	3.258	ug/l #	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093021\
Data File : VN068772.D
Acq On : 30 Sep 2021 19:11
Operator : JC/MD
Sample : M3915-12
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW17811-20210921

Quant Time: Oct 01 04:49:20 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092921W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 29 23:31:35 2021
Response via : Initial Calibration

