

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093022\
 Data File : VN074661.D
 Acq On : 30 Sep 2022 12:20
 Operator : JC\MD
 Sample : VN0930MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0930MBL01

Quant Time: Oct 03 02:04:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092922W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 30 07:44:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.928	168	365748	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.828	114	641758	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.616	117	634157	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.551	152	307138	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.287	65	238042	25.969	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	51.940%#
35) Dibromofluoromethane	7.869	113	235765	34.709	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	69.420%#
50) Toluene-d8	10.310	98	631807	22.113	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	44.220%#
62) 4-Bromofluorobenzene	12.604	95	310504	34.272	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	68.540%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093022\
Data File : VN074661.D
Acq On : 30 Sep 2022 12:20
Operator : JC\MD
Sample : VN0930MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0930MBL01

Quant Time: Oct 03 02:04:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092922W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 30 07:44:59 2022
Response via : Initial Calibration

