

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093022\
 Data File : VN074665.D
 Acq On : 30 Sep 2022 14:28
 Operator : JC\MD
 Sample : N4863-06DL 5X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-7DL

Quant Time: Oct 03 02:05:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092922W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 30 07:44:59 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.934	168	553507	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.828	114	1084339	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.616	117	1103723	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.551	152	490168	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.286	65	717372	51.713	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.420%
35) Dibromofluoromethane	7.863	113	604035	52.629	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.260%
50) Toluene-d8	10.310	98	2324989	48.160	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.320%
62) 4-Bromofluorobenzene	12.610	95	760952	49.710	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.420%
Target Compounds						Qvalue
11) Tert butyl alcohol	5.169	59	666181	179.819	ug/l	99
19) Methyl tert-butyl Ether	5.422	73	1342806	34.034	ug/l	100
22) Diisopropyl ether	6.298	45	34434	0.758	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093022\
Data File : VN074665.D
Acq On : 30 Sep 2022 14:28
Operator : JC\MD
Sample : N4863-06DL 5X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-7DL

Quant Time: Oct 03 02:05:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092922W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 30 07:44:59 2022
Response via : Initial Calibration

