

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093022\
 Data File : VN074666.D
 Acq On : 30 Sep 2022 14:52
 Operator : JC\MD
 Sample : N4863-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-2R

Quant Time: Oct 03 02:05:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092922W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 30 07:44:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.934	168	579785	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.828	114	1118247	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.622	117	1183005	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.551	152	562728	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.287	65	723715	49.806	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.620%
35) Dibromofluoromethane	7.869	113	598672	50.580	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.160%
50) Toluene-d8	10.310	98	2342283	47.047	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.100%
62) 4-Bromofluorobenzene	12.604	95	792743	50.216	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.440%
Target Compounds						
					Qvalue	
19) Methyl tert-butyl Ether	5.428	73	97631	2.362	ug/l	96
25) 2-Butanone	7.157	43	14567	1.089	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093022\
Data File : VN074666.D
Acq On : 30 Sep 2022 14:52
Operator : JC\MD
Sample : N4863-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-2R

Quant Time: Oct 03 02:05:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092922W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 30 07:44:59 2022
Response via : Initial Calibration

