

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093022\
 Data File : VN074668.D
 Acq On : 30 Sep 2022 15:40
 Operator : JC\MD
 Sample : N4913-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-01-092922

Quant Time: Oct 03 02:06:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092922W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 30 07:44:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.933	168	615240	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.827	114	1186785	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.616	117	1237327	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.545	152	580828	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.286	65	772621	50.107	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	100.220%
35) Dibromofluoromethane	7.869	113	629975	50.151	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.300%
50) Toluene-d8	10.310	98	2419281	45.787	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.580%
62) 4-Bromofluorobenzene	12.610	95	820345	48.963	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.920%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093022\
Data File : VN074668.D
Acq On : 30 Sep 2022 15:40
Operator : JC\MD
Sample : N4913-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-01-092922

Quant Time: Oct 03 02:06:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092922W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 30 07:44:59 2022
Response via : Initial Calibration

