

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093022\
 Data File : VN074671.D
 Acq On : 30 Sep 2022 16:53
 Operator : JC\MD
 Sample : N4913-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 GWOW-BR-02-182-202-092922

Quant Time: Oct 03 02:06:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092922W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 30 07:44:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.933	168	581805	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.827	114	1132269	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.616	117	1151034	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.551	152	521718	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.292	65	736826	50.532	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.060%
35) Dibromofluoromethane	7.869	113	605323	50.509	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.020%
50) Toluene-d8	10.310	98	2355824	46.733	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.460%
62) 4-Bromofluorobenzene	12.604	95	784062	49.051	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.100%
Target Compounds						
52) Toluene	10.374	92	40369	1.130	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093022\
Data File : VN074671.D
Acq On : 30 Sep 2022 16:53
Operator : JC\MD
Sample : N4913-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
GWOW-BR-02-182-202-092922

Quant Time: Oct 03 02:06:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092922W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 30 07:44:59 2022
Response via : Initial Calibration

