

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093022\
 Data File : VN074674.D
 Acq On : 30 Sep 2022 18:05
 Operator : JC\MD
 Sample : N4933-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-RW5-DUP-20220927

Quant Time: Oct 03 02:07:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092922W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 30 07:44:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.933	168	586446	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.828	114	1139687	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.616	117	1166071	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.551	152	533463	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.292	65	756708	51.485	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 102.980%			
35) Dibromofluoromethane	7.869	113	626190	51.910	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 103.820%			
50) Toluene-d8	10.310	98	2448379	48.252	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 96.500%			
62) 4-Bromofluorobenzene	12.610	95	800316	49.742	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 99.480%			
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	4.051	101	207976	19.246	ug/l	91
12) 1,1-Dichloroethene	4.028	96	153556	14.020	ug/l	95
27) cis-1,2-Dichloroethene	7.163	96	21205	1.449	ug/l	93
30) Chloroform	7.657	83	16769	0.741	ug/l #	83
38) Carbon Tetrachloride	8.033	117	15447	0.873	ug/l #	79
42) 1,2-Dichloroethane	8.386	62	7502	0.375	ug/l	94
44) Trichloroethene	9.075	130	3320833	267.693	ug/l	89
55) 1,1,2-Trichloroethane	10.774	97	11855	0.818	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093022\
Data File : VN074674.D
Acq On : 30 Sep 2022 18:05
Operator : JC\MD
Sample : N4933-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-RW5-DUP-20220927

Quant Time: Oct 03 02:07:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092922W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 30 07:44:59 2022
Response via : Initial Calibration

