

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093022\
 Data File : VN074675.D
 Acq On : 30 Sep 2022 18:29
 Operator : JC\MD
 Sample : N4933-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-RW5-GW-658-660

Quant Time: Oct 03 02:07:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092922W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 30 07:44:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.933	168	567237	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.827	114	1060629	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.621	117	1120977	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.551	152	514585	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.286	65	696288	48.978	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.960%
35) Dibromofluoromethane	7.869	113	581778	51.823	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.640%
50) Toluene-d8	10.310	98	2304075	48.793	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.580%
62) 4-Bromofluorobenzene	12.604	95	769600	51.398	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.800%

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.004	85	4487	0.378	ug/l #	60
9) 1,1,2-Trichlorotrifluo...	4.051	101	219599	21.010	ug/l	88
12) 1,1-Dichloroethene	4.028	96	147217	13.896	ug/l	95
24) 1,1-Dichloroethane	6.198	63	38753	1.631	ug/l #	92
27) cis-1,2-Dichloroethene	7.157	96	22022	1.556	ug/l	85
30) Chloroform	7.669	83	20285	0.927	ug/l #	71
38) Carbon Tetrachloride	8.039	117	17679	1.073	ug/l #	54
44) Trichloroethene	9.075	130	5753090	498.327	ug/l	91
55) 1,1,2-Trichloroethane	10.774	97	12999	0.964	ug/l #	85
64) Tetrachloroethene	10.845	164	6469	0.683	ug/l #	65

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093022\
Data File : VN074675.D
Acq On : 30 Sep 2022 18:29
Operator : JC\MD
Sample : N4933-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-RW5-GW-658-660

Quant Time: Oct 03 02:07:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092922W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 30 07:44:59 2022
Response via : Initial Calibration

