

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093022\
 Data File : VN074676.D
 Acq On : 30 Sep 2022 18:53
 Operator : JC\MD
 Sample : N4933-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-RW5-GW-678-680

Quant Time: Oct 03 02:07:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092922W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 30 07:44:59 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.934	168	557632	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.828	114	1075149	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.622	117	1133594	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.551	152	505472	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.287	65	758089	54.244	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.480%	
35) Dibromofluoromethane	7.863	113	623401	54.781	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 109.560%	
50) Toluene-d8	10.310	98	2484504	51.904	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.800%	
62) 4-Bromofluorobenzene	12.604	95	812240	53.514	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 107.020%	
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	4.051	101	187885	18.285	ug/l	Qvalue 90
12) 1,1-Dichloroethene	4.028	96	145018	13.924	ug/l	96
24) 1,1-Dichloroethane	6.198	63	44842	1.920	ug/l	97
25) 2-Butanone	7.175	43	16469	1.280	ug/l #	64
27) cis-1,2-Dichloroethene	7.163	96	19491	1.401	ug/l	99
30) Chloroform	7.657	83	18339	0.852	ug/l	87
32) 1,1,1-Trichloroethane	7.863	97	20290	1.074	ug/l	95
38) Carbon Tetrachloride	8.051	117	13260	0.794	ug/l #	96
44) Trichloroethene	9.075	130	3097663	264.693	ug/l	86
55) 1,1,2-Trichloroethane	10.769	97	12548	0.918	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093022\
Data File : VN074676.D
Acq On : 30 Sep 2022 18:53
Operator : JC\MD
Sample : N4933-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-RW5-GW-678-680

Quant Time: Oct 03 02:07:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092922W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 30 07:44:59 2022
Response via : Initial Calibration

