

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN093022\  
 Data File : VN074678.D  
 Acq On : 30 Sep 2022 19:42  
 Operator : JC\MD  
 Sample : N4860-01RE  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 20 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 MW-01RE

Quant Time: Oct 03 02:08:28 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N092922W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Sep 30 07:44:59 2022  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units  | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|--------|----------|
| -----                       |        |       |          |          |        |          |
| Internal Standards          |        |       |          |          |        |          |
| 1) Pentafluorobenzene       | 7.940  | 168   | 605353   | 50.000   | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene     | 8.828  | 114   | 1182379  | 50.000   | ug/l   | 0.00     |
| 63) Chlorobenzene-d5        | 11.616 | 117   | 1199600  | 50.000   | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 13.551 | 152   | 548903   | 50.000   | ug/l   | 0.00     |
| System Monitoring Compounds |        |       |          |          |        |          |
| 33) 1,2-Dichloroethane-d4   | 8.287  | 65    | 738925   | 48.705   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 74 - 125 | Recovery | =      | 97.400%  |
| 35) Dibromofluoromethane    | 7.869  | 113   | 605957   | 48.419   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =      | 96.840%  |
| 50) Toluene-d8              | 10.310 | 98    | 2344264  | 44.532   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 86 - 113 | Recovery | =      | 89.060%  |
| 62) 4-Bromofluorobenzene    | 12.610 | 95    | 770642   | 46.168   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | =      | 92.340%  |
| Target Compounds            |        |       |          |          |        |          |
|                             |        |       |          |          | Qvalue |          |
| 44) Trichloroethene         | 9.075  | 130   | 32394    | 2.517    | ug/l   | 90       |
| -----                       |        |       |          |          |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN093022\  
Data File : VN074678.D  
Acq On : 30 Sep 2022 19:42  
Operator : JC\MD  
Sample : N4860-01RE  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 20 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
MW-01RE

Quant Time: Oct 03 02:08:28 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N092922W.M  
Quant Title : SW846 8260  
QLast Update : Fri Sep 30 07:44:59 2022  
Response via : Initial Calibration

