

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093022\
 Data File : VN074683.D
 Acq On : 01 Oct 2022 02:22
 Operator : JC\MD
 Sample : VN0930WBL02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0930WBL02

Quant Time: Oct 03 02:09:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092922W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 30 07:44:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.934	168	560548	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.828	114	1054333	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.622	117	1065009	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.551	152	502397	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.286	65	689953	49.112	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.220%
35) Dibromofluoromethane	7.869	113	570079	51.084	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.160%
50) Toluene-d8	10.310	98	2158234	45.978	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.960%
62) 4-Bromofluorobenzene	12.610	95	713231	47.918	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.840%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093022\
Data File : VN074683.D
Acq On : 01 Oct 2022 02:22
Operator : JC\MD
Sample : VN0930WBL02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0930WBL02

Quant Time: Oct 03 02:09:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092922W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 30 07:44:59 2022
Response via : Initial Calibration

