

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093022\
 Data File : VN074691.D
 Acq On : 01 Oct 2022 05:37
 Operator : JC\MD
 Sample : N4882-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-105

Quant Time: Oct 03 02:11:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092922W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 30 07:44:59 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.934	168	504619	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.828	114	996340	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.622	117	1025441	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.551	152	467875	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.286	65	650440	51.431	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.860%
35) Dibromofluoromethane	7.869	113	543744	51.560	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.120%
50) Toluene-d8	10.310	98	2005459	45.210	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.420%
62) 4-Bromofluorobenzene	12.610	95	666144	47.360	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.720%
Target Compounds						
						Qvalue
20) Methylene Chloride	4.893	84	70281	4.990	ug/l	87
30) Chloroform	7.669	83	19971	1.025	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093022\
Data File : VN074691.D
Acq On : 01 Oct 2022 05:37
Operator : JC\MD
Sample : N4882-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-105

Quant Time: Oct 03 02:11:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092922W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 30 07:44:59 2022
Response via : Initial Calibration

