

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093022\
 Data File : VN074698.D
 Acq On : 01 Oct 2022 08:27
 Operator : JC\MD
 Sample : N4882-10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-119

Quant Time: Oct 03 02:13:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092922W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 30 07:44:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.939	168	503992	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.828	114	948083	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.616	117	936868	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.551	152	423150	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.292	65	640484	50.707	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.420%
35) Dibromofluoromethane	7.863	113	522480	52.066	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.140%
50) Toluene-d8	10.316	98	1981013	46.932	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.860%
62) 4-Bromofluorobenzene	12.604	95	642976	48.039	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.080%
Target Compounds						
19) Methyl tert-butyl Ether	5.434	73	29377	0.818	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093022\
Data File : VN074698.D
Acq On : 01 Oct 2022 08:27
Operator : JC\MD
Sample : N4882-10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-119

Quant Time: Oct 03 02:13:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092922W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 30 07:44:59 2022
Response via : Initial Calibration

