

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093022\
 Data File : VN074703.D
 Acq On : 01 Oct 2022 10:28
 Operator : JC\MD
 Sample : N4882-16
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-128

Quant Time: Oct 03 02:14:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092922W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 30 07:44:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.933	168	494957	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.828	114	939716	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.616	117	942808	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.551	152	426997	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.286	65	641684	51.729	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.460%
35) Dibromofluoromethane	7.869	113	526189	52.902	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.800%
50) Toluene-d8	10.310	98	1988767	47.535	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.080%
62) 4-Bromofluorobenzene	12.604	95	632979	47.713	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.420%
Target Compounds						
						Qvalue
16) Acetone	4.134	43	60009	4.373	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093022\
Data File : VN074703.D
Acq On : 01 Oct 2022 10:28
Operator : JC\MD
Sample : N4882-16
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-128

Quant Time: Oct 03 02:14:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092922W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 30 07:44:59 2022
Response via : Initial Calibration

